



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AZZUR TECHNICAL SERVICES
465 Waverly Oaks Rd, Suite 217
Waltham, MA 02452
Miguel Menendez Phone: 781 893 1251

CALIBRATION

Valid To: May 31, 2023

Certificate Number: 6129.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 5}:

I. Mechanical

Parameter/Equipment	Range	CMC ^{2, 4} (±)	Comments
Pressure – Measuring Equipment ³	Differential Pressure		
	(-1 to 1) inH2O (-5 to 5) inH2O	0.0025 inH2O 0.0079 inH2O	Additel ADT760 with ADT155
Gauge Pressure	(0 to 10) inH2O	0.021 inH2O	Fluke 750P01
	(-13.5 to 100) psi	0.039 psi	Additel ADT761A with ADT155
	(100 to 300) psi (300 to 500) psi	0.070 psi 0.12 psi	

II. Thermodynamics

Parameter/Equipment	Range	CMC ^{2, 4} (±)	Comments
Temperature – Measuring Equipment ³	(-30 to -25) °C	0.070 °C	PRT with fluid bath or drywell
	(-25 to 140) °C	0.044 °C	
	(140 to 200) °C	0.055 °C	

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
Temperature – Measure ³	(-196 to -37) °C (-37 to 0) °C (0 to 200) °C (200 to 420) °C	0.031 °C 0.022 °C 0.026 °C 0.043 °C	PRT with digital readout

¹ This laboratory offers commercial and field calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g. resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.

⁴ The type of instrument or material being calibrated is defined by the parameter. This indicates the laboratory is capable of calibrating instruments that measure or generate the values in the ranges indicated for the listed measurement parameter.

⁵ This scope meets A2LA's P112 *Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

AZZUR TECHNICAL SERVICES

Waltham, MA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – *Specific Requirements: Calibration Laboratory Accreditation Program*. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 19th day of April 2021.

A blue ink signature of the Vice President of Accreditation Services.

Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 6129.01
Valid to May 31, 2023

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.